

web: [www.cacgas.com.au](http://www.cacgas.com.au)  
 email: [cac@cacgas.com.au](mailto:cac@cacgas.com.au)  
 phone: 1300 CAC GAS fax: (02) 8676 6599

# CAC

GAS & INSTRUMENTATION

## Certificate of Composition

**Part No. 112DA4GASSP10N**

**Batch No. 13152-3-2**

**Cylinder No C198426**

**Cylinder Valve C10**

**Cylinder Size 112DA**

**Nett Weight (kg) 0.13**

Component	Requested Value	Certified Value
Hydrogen Sulphide	25ppm	25ppm
Carbon Monoxide	100ppm	100.2ppm
Propane	1.1%	1.101%
Oxygen	18%	17.98%
Nitrogen	Balance	Balance

**Pressure 1000 psi      Volume 112L**

**Certified by: M. Stephenson**

**Manufactured Date: 15-05-2024**

**Valid until: 15-05-2026**

UN Number: 1956      Class 2.2

Compressed Gas, N.O.S, (Oxygen, Nitrogen Mixture)

Please note all units are in mol% and mixture accuracy +/-2%. Relative mixtures are traceable to standards calibrated at the National Physical Laboratory, Teddington, Middlesex, England. The UK National Physical Laboratory standards are internationally recognised and directly equivalent to the USA NIST Standards.

